

Certificate of calibration

Indicated	Etalon	Correction
1001m	1000m	-1m
1993m	2000m	+7m
3001m	3000m	-1m
4000m	4000m	0m
4997m	5000m	+3m
6000m	6000m	0m
6993m	7000m	+7m
7991m	8000m	+9m
6996m	7000m	+4m
5996m	6000m	+4m
4998m	5000m	+2m
3997m	4000m	+3m
2999m	3000m	+1m
1996m	2000m	+4m
998m	1000m	+2m

LXN
P. Mihelin
PI

